synergyhealth
our work protects your world

Synergy Sterilisation Rawang (M) Sdn Bhd (234258-U)
LOT 42, JALAN INDUSTRI 2/1, RAWANG INTEGRATED INDUSTRIAL PARK,
48000 RAWANG, SELANGOR, MALAYSIA.
Tel: +60(3) 6099 9600 Fax: +60(3) 6091 5306
Website: https://www.synergyhealthplc.com

Certificate No: **13073845**

CERTIFICATE OF IRRADIATION

CUSTOMER : NITRITEX (M) SDN BHD

ADDRESS : No:2, Jalan Jurunilai U1/20,Seksyen U1, Hicom Glenmarie Industrial Park, 40150 Shah Alam, Selangor Darul Ehsan.

DOSE REQUESTED : MINIMUM : 25 kGy MAXIMUM : 40 kGy

IRRADIATION DATE : 27.07.2013 **IRRADIATION LOT NO.** : 13073845

PRODUCT DESCRIPTION : Bioclean Extra 400 St Latex (BLAS)

Manufacturing Lot No.	Batch No.	Product Code / Size	Quantity (Cartons)
13-1143	L61307020	BLAS-M	10
13-1072	L61307007	BLAS-M	04
13-1148	L61307023	BLAS-L	02
		Total	**16**

has been successfully irradiated and received the dose specification requested (as the above stated) within the precision and accuracy of dosimetry system employed

Min. Dose Delivered (kGy) : **25.4 kGy** **Max. Dose Delivered (kGy)** : **35.8 kGy**

Approved By : 29/07/2013

Noraini binti Sudin
Senior Executive QA
Synergy Sterilisation Rawang (M) Sdn. Bhd.
(234258-U)