



Synergy Sterilisation Rawang (M) Sdn Bhd (234258-U)  
LOT 42, JALAN INDUSTRI 2/1, RAWANG INTEGRATED INDUSTRIAL PARK,  
48000 RAWANG, SELANGOR, MALAYSIA.  
Tel: +60(3) 6099 9600 Fax: +60(3) 6091 5306  
Website: <https://www.synergyhealthplc.com>

Certificate No: **13116029**

## CERTIFICATE OF IRRADIATION

CUSTOMER : NITRITEX (M) SDN BHD

ADDRESS : No:2, Jalan Jurunilai U1/20, Seksyen U1, Hicom Glenmarie Industrial Park,  
40150 Shah Alam, Selangor Darul Ehsan.

DOSE REQUESTED : MINIMUM : 25 kGy MAXIMUM : 40 kGy

IRRADIATION DATE : 20.11.2013 IRRADIATION LOT NO. : 13116029

PRODUCT DESCRIPTION : Bioclean Extra 400 St Latex (BLAS)

Manufacturing Lot No.	Batch No.	Product Code / Size	Quantity (Cartons)
Packing:- 200 Pairs/ctn			
13-1867	L61311009	BLAS-M	80
Total			80

has been successfully irradiated and received the dose specification requested (as the above stated)  
within the precision and accuracy of dosimetry system employed

Min. Dose Delivered (kGy) : 25.7 kGy Max. Dose Delivered (kGy) : 32.5 kGy

Approved By

*Noraini* 21/11/2013  
Noraini binti Sudin  
Senior Executive QA  
Synergy Sterilisation Rawang (M) Sdn. Bhd,  
(234258-U)