synergyhealth
our work protects your world

Synergy Sterilisation Rawang (M) Sdn Bhd (234258-U)
LOT 42, JALAN INDUSTRI 2/1, RAWANG INTEGRATED INDUSTRIAL PARK,
48000 RAWANG, SELANGOR, MALAYSIA.
Tel: +60(3) 6099 9600 Fax: +60(3) 6091 5306
Website: https://www.synergyhealthplc.com

Certificate No: **13041836**

CERTIFICATE OF IRRADIATION

CUSTOMER	:	NITRITEX (M) SDN BHD
ADDRESS	:	No:2, Jalan Jurunilai U1/20,Seksyen U1, Hicom Glenmarie Industrial Park, 40150 Shah Alam, Selangor Darul Ehsan.

DOSE REQUESTED	:	MINIMUM	:	25 kGy	MAXIMUM	:	40 kGy
IRRADIATION DATE	:	12.04.2013		**IRRADIATION LOT NO.**		:	13041836
PRODUCT DESCRIPTION	:	Bioclean Extra 400 St Latex (BLAS)					

Manufacturing Lot No.	Batch No.	Product Code / Size	Quantity (Cartons)
Packing:- 200 Pairs/ctn			
13-0515	L61303038	BLAS-S	32
		Total	**32**

has been successfully irradiated and received the dose specification requested (as the above stated) within the precision and accuracy of dosimetry system employed

Min. Dose Delivered (kGy) : **27.0 kGy** **Max. Dose Delivered (kGy)** : **38.2 kGy**

Approved By : Noraini 12/04/2013

Noraini binti Sudin
Senior Executive QA
Synergy Sterilisation Rawang (M) Sdn. Bhd
(234258-U)