synergyhealth

our work protects your world

www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 03-Jan-2014 MY01S11092899-1-1

This is to certify that Synergy Sterilisation Rawang (M) Sdn. Bhd. has where appropriate delivered an irradiation process
in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

NITRITEX (M) SDN BHD
No.2, Jalan Jurunilai U1/20
Seksyen U1, Hicom Glenmarie
Industrial Park
40150 Shah Alam
Selangor

Order Information

Account Number:	101178
Synergy Health Sales Part Reference:	1055043
Customer Reference Number:	SNB13/0263
Product Description:	Bioclean Extra 400 St Latex
Validation Reference:	0.0171
Quantity Received:	64
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	40.0
Customer Unit Lot/Batch Number:	SNB13/0263
Other Process Details:	Product Description: Bioclean Extra 400 St Latex (BLAS) Packing: 200 Pairs/ctn Mfg Lot No Batch No Pro. Code/Size Qty (Ctns) 13-2070 L61312019 BLAS-L 64 Total: 64 Cartons

Irradiation Data

Date and Time of Irradiation:	03-Jan-2014 02:31
Reference Dose Range kGy:	31.0 - 31.0
Calculated Minimum Dose kGy:	25.7
Calculated Maximum Dose kGy:	31.8

Irradiation Release Authorised By Synergy Health plc