


<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 05-Jul-2014

UK33S11193276-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
 Minton Enterprise Park
 Oaks Drive
 Newmarket
 Suffolk
 CB8 7YY

Order Information

Account Number:	124123												
Synergy Health Sales Part Reference:	1073645												
Customer Reference Number:	0000006444												
Product Description:	Latex Gloves Group 3 DV4185 25-45kGy												
Validation Reference:	4185												
Quantity Received:	36												
Customer Minimum Specification kGy:	25.0												
Customer Maximum Specification kGy:	45.0												
Customer Unit Lot/Batch Number:	0000006444												
Other Process Details:	Irradiation No: S1119327611 Description: Bioclean Extra 400 St Latex (BLAS) Packing: 200 Pairs /ctn												
	<table><tr><td>Lot No</td><td>Batch No</td><td>Product Code</td><td>Qty(ctns)</td></tr><tr><td></td><td></td><td>/ Size</td><td></td></tr><tr><td>14-0748</td><td>L61404036</td><td>BLAS-S</td><td>36</td></tr></table>	Lot No	Batch No	Product Code	Qty(ctns)			/ Size		14-0748	L61404036	BLAS-S	36
Lot No	Batch No	Product Code	Qty(ctns)										
		/ Size											
14-0748	L61404036	BLAS-S	36										
	Total: 36 Cartons												

Irradiation Data

Date and Time of Irradiation:	05-Jul-2014 02:59
Reference Dose Range kGy:	34.2 - 34.9
Calculated Minimum Dose kGy:	30.5
Calculated Maximum Dose kGy:	37.5

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069