


<http://www.synergyhealthplc.com>

# Certificate of Irradiation

Date Issued: 02-Aug-2014

UK33S11208554-3-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products  
 EN ISO 9001:2008 Quality Management System  
 EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd  
 Minton Enterprise Park  
 Oaks Drive  
 Newmarket  
 Suffolk  
 CB8 7YY

## Order Information

Account Number:	124123												
Synergy Health Sales Part Reference:	1073645												
Customer Reference Number:	0000006538												
Product Description:	Latex Gloves Group 3 DV4185 25-45kGy												
Validation Reference:	4185												
Quantity Received:	48												
Customer Minimum Specification kGy:	25.0												
Customer Maximum Specification kGy:	45.0												
Customer Unit Lot/Batch Number:	0000006538												
Other Process Details:	Irradiation No: S1120855431 Description: Bioclean Extra 400 St Latex (BLAS) Packing: 200 Pairs /ctn												
	<table><tr><td>Lot No</td><td>Batch No</td><td>Product Code</td><td>Qty(ctns)</td></tr><tr><td></td><td></td><td>/ Size</td><td></td></tr><tr><td>14-0944</td><td>L61405046</td><td>BLAS-M</td><td>48</td></tr></table>	Lot No	Batch No	Product Code	Qty(ctns)			/ Size		14-0944	L61405046	BLAS-M	48
Lot No	Batch No	Product Code	Qty(ctns)										
		/ Size											
14-0944	L61405046	BLAS-M	48										
	Total: 48 Cartons												

## Irradiation Data

Date and Time of Irradiation:	02-Aug-2014 06:28
Reference Dose Range kGy:	35.3 - 35.6
Calculated Minimum Dose kGy:	31.5
Calculated Maximum Dose kGy:	38.2

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM  
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069