



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 26-Aug-2014

UK33S11219983-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
Minton Enterprise Park
Oaks Drive
Newmarket
Suffolk
CB8 7YY

Order Information

Account Number:	124123
Synergy Health Sales Part Reference:	1073645
Customer Reference Number:	0000006582
Product Description:	Latex Gloves Group 3 DV4185 25-45kGy
Validation Reference:	4185
Quantity Received:	95
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	0000006582


<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 26-Aug-2014

UK33S11219983-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
 Minton Enterprise Park
 Oaks Drive
 Newmarket
 Suffolk
 CB8 7YY

Other Process Details:

Irradiation No: S1121998321

Description: Bioclean Extra 400 St Latex (BLAS)

Packing: 200 Pairs /ctn

Lot No	Batch No	Product Code	Qty(ctns) / Size
14-1065	L61406018	BLAS-M	48

Description: Bioclean Extra 400 St Latex (BLAS)

Packing: 200 Pairs /ctn

Lot No	Batch No	Product Code	Qty(ctns) / Size
14-0936	L61405042	BLAS-S	12

Description: Bioclean Extra 400 St Latex (BLAS)

Packing: 200 Pairs /ctn

Lot No	Batch No	Product Code	Qty(ctns) / Size
14-1052	L61406016	BLAS-S	35

Total: 95 Cartons

Irradiation Data

Date and Time of Irradiation:	23-Aug-2014 20:09
Reference Dose Range kGy:	32.8 - 33.0
Calculated Minimum Dose kGy:	29.3
Calculated Maximum Dose kGy:	35.5



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 26-Aug-2014

UK33S11219983-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
Minton Enterprise Park
Oaks Drive
Newmarket
Suffolk
CB8 7YY

Irradiation Release Authorised By Synergy Health plc