


<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 15-Sep-2014

UK33S11235179-3-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
 Minton Enterprise Park
 Oaks Drive
 Newmarket
 Suffolk
 CB8 7YY

Order Information

Account Number:	124123												
Synergy Health Sales Part Reference:	1073645												
Customer Reference Number:	0000006603												
Product Description:	Latex Gloves Group 3 DV4185 25-45kGy												
Validation Reference:	4185												
Quantity Received:	80												
Customer Minimum Specification kGy:	25.0												
Customer Maximum Specification kGy:	45.0												
Customer Unit Lot/Batch Number:	0000006603, 2plts												
Other Process Details:	Irradiation No: S1123517931 Description: Bioclean Extra 400 St Latex (BLAS) Packing: 200 Pairs /ctn												
	<table><tr><td>Lot No</td><td>Batch No</td><td>Product Code</td><td>Qty(ctns)</td></tr><tr><td></td><td></td><td>/ Size</td><td></td></tr><tr><td>14-1315</td><td>L61407034</td><td>BLAS-S</td><td>32</td></tr></table>	Lot No	Batch No	Product Code	Qty(ctns)			/ Size		14-1315	L61407034	BLAS-S	32
Lot No	Batch No	Product Code	Qty(ctns)										
		/ Size											
14-1315	L61407034	BLAS-S	32										
	Description: Bioclean Extra 400 St Latex (BLAS) Packing: 200 Pairs /ctn												
	<table><tr><td>Lot No</td><td>Batch No</td><td>Product Code</td><td>Qty(ctns)</td></tr><tr><td></td><td></td><td>/ Size</td><td></td></tr><tr><td>14-1271</td><td>L61407026</td><td>BLAS-XL</td><td>48</td></tr></table>	Lot No	Batch No	Product Code	Qty(ctns)			/ Size		14-1271	L61407026	BLAS-XL	48
Lot No	Batch No	Product Code	Qty(ctns)										
		/ Size											
14-1271	L61407026	BLAS-XL	48										
	Total: 80 Cartons												

Irradiation Data

Date and Time of Irradiation:	12-Sep-2014 12:44
-------------------------------	-------------------



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 15-Sep-2014

UK33S11235179-3-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
Minton Enterprise Park
Oaks Drive
Newmarket
Suffolk
CB8 7YY

Reference Dose Range kGy:	32.3 - 33.5
Calculated Minimum Dose kGy:	28.9
Calculated Maximum Dose kGy:	36.0

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069