


<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 13-Oct-2014

UK33S11252136-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
 Minton Enterprise Park
 Oaks Drive
 Newmarket
 Suffolk
 CB8 7YY

Order Information

Account Number:	124123												
Synergy Health Sales Part Reference:	1073645												
Customer Reference Number:	0000006699												
Product Description:	Latex Gloves Group 3 DV4185 25-45kGy												
Validation Reference:	4185												
Quantity Received:	36												
Customer Minimum Specification kGy:	25.0												
Customer Maximum Specification kGy:	45.0												
Customer Unit Lot/Batch Number:	0000006699												
Other Process Details:	Irradiation No: S1125213621 Description: Bioclean Extra 400 St Latex (BLAS) Packing: 200 Pairs /ctn												
	<table><tr><td>Lot No</td><td>Batch No</td><td>Product Code</td><td>Qty(ctns)</td></tr><tr><td></td><td></td><td>/ Size</td><td></td></tr><tr><td>14-1082</td><td>L61406020</td><td>BLAS-M</td><td>36</td></tr></table>	Lot No	Batch No	Product Code	Qty(ctns)			/ Size		14-1082	L61406020	BLAS-M	36
Lot No	Batch No	Product Code	Qty(ctns)										
		/ Size											
14-1082	L61406020	BLAS-M	36										
	Total: 36 Cartons												

Irradiation Data

Date and Time of Irradiation:	11-Oct-2014 20:44
Reference Dose Range kGy:	32.8 - 33.5
Calculated Minimum Dose kGy:	29.3
Calculated Maximum Dose kGy:	36.0

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069