



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 20-Oct-2014

UK33S11256293-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
Minton Enterprise Park
Oaks Drive
Newmarket
Suffolk
CB8 7YY

Order Information

Account Number:	124123
Synergy Health Sales Part Reference:	1073645
Customer Reference Number:	0000006700
Product Description:	Latex Gloves Group 3 DV4185 25-45kGy
Validation Reference:	4185
Quantity Received:	94
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	0000006700


<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 20-Oct-2014

UK33S11256293-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
 Minton Enterprise Park
 Oaks Drive
 Newmarket
 Suffolk
 CB8 7YY

Other Process Details:

Irradiation No: S1125629321

Description: Bioclean Extra 400 St Latex (BLAS)

Packing: 200 Pairs /ctn

Lot No	Batch No	Product Code	Qty(ctns) / Size
14-0900	L61405038	BLAS-L	22

Description: Bioclean Extra 400 St Latex (BLAS)

Packing: 200 Pairs /ctn

Lot No	Batch No	Product Code	Qty(ctns) / Size
14-0939	L61405045	BLAS-L	36

Description: Bioclean Extra 400 St Latex (BLAS)

Packing: 200 Pairs /ctn

Lot No	Batch No	Product Code	Qty(ctns) / Size
14-1082	L61406020	BLAS-M	36

Total: 94 Cartons

Irradiation Data

Date and Time of Irradiation:	18-Oct-2014 06:57
Reference Dose Range kGy:	33.3 - 34.0
Calculated Minimum Dose kGy:	29.7
Calculated Maximum Dose kGy:	36.5



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 20-Oct-2014

UK33S11256293-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
Minton Enterprise Park
Oaks Drive
Newmarket
Suffolk
CB8 7YY

Irradiation Release Authorised By Synergy Health plc