


<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 01-Nov-2014

UK33S11266841-7-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
 Minton Enterprise Park
 Oaks Drive
 Newmarket
 Suffolk
 CB8 7YY

Order Information

Account Number:	124123																				
Synergy Health Sales Part Reference:	1073645																				
Customer Reference Number:	0000006724																				
Product Description:	Latex Gloves Group 3 DV4185 25-45kGy																				
Validation Reference:	4185																				
Quantity Received:	87																				
Customer Minimum Specification kGy:	25.0																				
Customer Maximum Specification kGy:	45.0																				
Customer Unit Lot/Batch Number:	0000006724																				
Other Process Details:	Irradiation No: S1126684171 Description: Bioclean Extra 400 St Latex (BLAS) Packing: 200 Pairs /ctn																				
	<table><tr><th>Lot No</th><th>Batch No</th><th>Product Code</th><th>Qty(ctns)</th></tr><tr><td></td><td></td><td>/ Size</td><td></td></tr><tr><td>14-0939</td><td>L61405045</td><td>BLAS-L</td><td>35</td></tr><tr><td>14-1032</td><td>L61406012</td><td>BLAS-L</td><td>50</td></tr><tr><td>14-1403</td><td>L61408004</td><td>BLAS-L</td><td>2</td></tr></table>	Lot No	Batch No	Product Code	Qty(ctns)			/ Size		14-0939	L61405045	BLAS-L	35	14-1032	L61406012	BLAS-L	50	14-1403	L61408004	BLAS-L	2
Lot No	Batch No	Product Code	Qty(ctns)																		
		/ Size																			
14-0939	L61405045	BLAS-L	35																		
14-1032	L61406012	BLAS-L	50																		
14-1403	L61408004	BLAS-L	2																		
	Total: 87 Cartons																				

Irradiation Data

Date and Time of Irradiation:	01-Nov-2014 10:35
Reference Dose Range kGy:	31.9 - 32.2
Calculated Minimum Dose kGy:	28.5
Calculated Maximum Dose kGy:	34.6



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 01-Nov-2014

UK33S11266841-7-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
Minton Enterprise Park
Oaks Drive
Newmarket
Suffolk
CB8 7YY

Irradiation Release Authorised By Synergy Health plc