



www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 10-Nov-2014

MY03S11271578-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

NITRITEX (M) SDN BHD
 No.2, Jalan Jurunilai U1/20
 Seksyen U1, Hicom Glenmarie
 Industrial Park
 40150 Shah Alam
 Selangor

Order Information

Account Number:	101178												
Synergy Health Sales Part Reference:	1072156												
Customer Reference Number:	SNB14/0143												
Product Description:	BIOCLEAN EXTRA 400 ST LATEX (BLAS)												
Quantity Received:	71												
Customer Minimum Specification kGy:	25.0												
Customer Maximum Specification kGy:	40.0												
Customer Unit Lot/Batch Number:	.												
Other Process Details:	DESCRIPTION: BIOCLEAN EXTRA 400 ST LATEX (BLAS) PACKING : 200 PAIR/CTN <table><tr><th>PRODUCT CODE</th><th>LOT NO</th><th>BATCH NO</th><th>QTY</th></tr><tr><td>BLAS-S</td><td>14-1946</td><td>L61410023</td><td>64</td></tr><tr><td>BLAS-XL</td><td>14-1271</td><td>L61407026</td><td>7</td></tr></table> ACTUAL DOSE RECEIVED: MIN:27.8 kGy MAX:33.5 kGy	PRODUCT CODE	LOT NO	BATCH NO	QTY	BLAS-S	14-1946	L61410023	64	BLAS-XL	14-1271	L61407026	7
PRODUCT CODE	LOT NO	BATCH NO	QTY										
BLAS-S	14-1946	L61410023	64										
BLAS-XL	14-1271	L61407026	7										

Irradiation Data

Date and Time of Irradiation:	07-Nov-2014 11:21
-------------------------------	-------------------

Registered Office: Suite 18.01 18th Floor, MWE Plaza, 8, Lebuh Farquhar, 10200 , Penang , MALAYSIA

VAT Number:



www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 10-Nov-2014

MY03S11271578-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

NITRITEK (M) SDN BHD
No.2, Jalan Jurunilai U1/20
Seksyen U1, Hicom Glenmarie
Industrial Park
40150 Shah Alam
Selangor

Irradiation Release Authorised By Synergy Health plc

Registered Office: Suite 18.01 18th Floor, MWE Plaza, 8, Lebuhr Farquhar, 10200, Penang, MALAYSIA

VAT Number:

Page 2 of 2