

http://www.synergyhealthplc.com

## **Certificate of Irradiation**

Date Issued: 15-Feb-2015 UK33S11326946-3-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

## **Order Information**

Account Number:	124123
Synergy Health Sales Part Reference:	1073645
Customer Reference Number:	000006921
Product Description:	Latex Gloves Group 3 DV4185 25-45kGy
Validation Reference:	4185
Quantity Received:	96
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	0000006921
Other Process Details:	Irradiation No: S1132694631
	Description: Bioclean Extra 400 St Latex (BLAS)
	Packing: 200 Pairs /ctn
	Lot No Batch No Product Code Qty(ctns) / Size
	14-2294 L61412004 BLAS-L 48
	14-2108 L61411002 BLAS-S 48
	Total: 96 Cartons

Irradiation Release Authorised By Synergy Health plc

Date and Time of Irradiation:

Reference Dose Range kGy:

Calculated Minimum Dose kGy: Calculated Maximum Dose kGy:

**Irradiation Data** 

15-Feb-2015 18:32

32.1 - 33.5 28.7

36.0