

DOCUMENT NO: QP-07-COI
REVISION NO: 6/02

STERILGAMMA (M) SDN. BHD. (234258-U)

LOT 42, RAWANG INTEGRATED INDUSTRIAL PARK,
48000 RAWANG, SELANGOR, MALAYSIA
TEL: +60(3) 6091 5301 (4 lines) FAX: +60(3) 6091 5320
Website: <http://www.sterilgamma.com>
email: feedback@sterilgamma.com

Certificate of Irradiation

This is to certify that

Customer: NITRITEX (M) SDN BHD
Address No: 2, Jalan Jurunilai U1/20, Seksyen U1, Hicom Glenmarie
Industrial Park, 40150 Shah Alam, Selangor Darul Ehsan.
Type of Product: Refer Attachment No: (4069)
Manufacturing Lot No: Refer Attachment No: (4069)
No. of Cartons: 126 Cartons
Requested Dose (min): 25.0 kGy



STERILGAMMA

From Our State-of-the-Art Sterilization & Logistics Plants

has been successfully irradiated and received the minimum dose requested (as the above stated) within the precision and accuracy of dosimetry system employed

Irradiation Lot No: 11104069
Irradiation Date: 10 October, 2011
Carton Serial No: 817487 - 817612
Dose Delivered (min): 26.7 kGy Max: 30.1 kGy

Certified By

Designation: Head Of Quality Assurance
Date: 12 October, 2011

DOCUMENT NO: QP-07-COI
REVISION NO: 6/02

STERILGAMMA (M) SDN. BHD. (234258-U)

LOT 42, RAWANG INTEGRATED INDUSTRIAL PARK,
48000 RAWANG, SELANGOR, MALAYSIA
TEL: +60(3) 6091 5301 (4 lines) FAX: +60(3) 6091 5320
Website: <http://www.sterilgamma.com>
email: feedback@sterilgamma.com

Attachment No : (4069)

Irradiation Lot No : 11104069

Page : 1 of 1

Product Code	Mfg Lot No./Batch No.	Size	Unit (Pairs)	No. Of Ctns
Bioclean Alpha Latex Gloves	11-1108			
	L31109025	(AL300-75)	7,000	35
	11-1109			
	L31109026	(AL300-90)	3,000	15
	1062/11 T			
	L31109012	(T-AL300-70)	5,600	28
Bioclean Advance Sterile Latex HS Gloves	11-1097			
	L11109021	(BASL-75)	9,600	48
Total			25,200	126

STERILGAMMA

Dear Sir, Please find enclosed the invoice for your reference.

Verify by

ab 

Production Manager
SterilGamma (M) Sdn Bhd
Date: December 29, 2011

Revision: 1
Date: 29/12/2011