

CERTIFICATE OF IRRADIATION

CUSTOMER : NITRITEX (M) SDN BHD

ADDRESS : No.2, Jalan Jurunilai U1/20, Seksyen U1, Hicom Glenmarie Industrial Park,
40150 Shah Alam, Selangor Darul Ehsan.

DOSE REQUESTED : MINIMUM : 25 kGy MAXIMUM : 40 kGy

IRRADIATION DATE : 09.09.2012 IRRADIATION LOT NO. : 12093934

PRODUCT DESCRIPTION : As Below

Lot No. / Batch No.	Batch No.	Product Code / Size	Quantity (Cartons)
Bioclean Alpha St Ltx (AL300) – Packing: 200 pairs/ctn			
12-1215	L31208035	AL300-80	22
12-1216	L31208036	AL300-85	34
Bioclean Alpha Tacky St Ltx (T-AL300) – Packing: 200 pairs/ctn			
12-1103-2	L31208002	T-AL300-B-7.5	04
1174/12T	L31208031	T-AL300-7.5	36
Bioclean Advance St Latex (BASL) – Packing: 200 pairs/ctn			
12-1207	L11208033	BASL8.0	02
12-1184	L11208040	BASL8.5	28
Total			126

has been successfully irradiated and received the dose specification requested (as the above stated)
within the precision and accuracy of dosimetry system employed

Min. Dose Delivered (kGy) : 26.9 kGy Max. Dose Delivered (kGy) : 29.9 kGy

Approved By

Noraini 10/09/2012
Noraini binti Sudin
Senior Executive Op
Synergy Sterilisation Rawang (M) Sdn. Bhd.
(234258-U)