synergyhealth

our work protects your world

www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 16-Feb-2014 MY01S11113139-1-1

This is to certify that Synergy Sterilisation Rawang (M) Sdn. Bhd. has where appropriate delivered an irradiation process
in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

NITRITEX (M) SDN BHD
No.2, Jalan Jurunilai U1/20
Seksyen U1, Hicom Glenmarie
Industrial Park
40150 Shah Alam
Selangor

Order Information

Account Number:	101178
Synergy Health Sales Part Reference:	1055042
Customer Reference Number:	SNB14/0042
Product Description:	Gloves
Validation Reference:	11/1(Nitritex)(03)
Quantity Received:	36
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	40.0
Customer Unit Lot/Batch Number:	SNB14/0042
Other Process Details:	Irradiation No: S1111313911 Description: Bioclean Alpha St Latex (AL300) Packing: 200 Pairs/ ctn

Lot No	Batch No	Product Code /Size	Qty(ctns)
14-0170	L31401036	AL300-75	36

Total : 36 Carton

Irradiation Data

Date and Time of Irradiation:	16-Feb-2014 04:00
Reference Dose Range kGy:	27.8 - 28.7
Calculated Minimum Dose kGy:	27.4
Calculated Maximum Dose kGy:	29.3

Irradiation Release Authorised By Synergy Health plc