


<http://www.synergyhealthplc.com>

# Certificate of Irradiation

Date Issued: 02-Aug-2014

UK33S11208554-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products  
 EN ISO 9001:2008 Quality Management System  
 EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd  
 Minton Enterprise Park  
 Oaks Drive  
 Newmarket  
 Suffolk  
 CB8 7YY

## Order Information

Account Number:	124123												
Synergy Health Sales Part Reference:	1073638												
Customer Reference Number:	0000006538												
Product Description:	Latex Gloves Group 1 DV4172 25-45kGy												
Validation Reference:	4172												
Quantity Received:	48												
Customer Minimum Specification kGy:	25.0												
Customer Maximum Specification kGy:	45.0												
Customer Unit Lot/Batch Number:	0000006538												
Other Process Details:	Irradiation No: S1120855411 Description: Bioclean Alpha St Latex (AL300) Packing: 200 Pairs /ctn												
	<table><tr><td>Lot No</td><td>Batch No</td><td>Product Code</td><td>Qty(ctns)</td></tr><tr><td></td><td></td><td>/ Size</td><td></td></tr><tr><td>14-1097</td><td>L31406025</td><td>AL300-75</td><td>48</td></tr></table>	Lot No	Batch No	Product Code	Qty(ctns)			/ Size		14-1097	L31406025	AL300-75	48
Lot No	Batch No	Product Code	Qty(ctns)										
		/ Size											
14-1097	L31406025	AL300-75	48										
	Total: 48 Cartons												

## Irradiation Data

Date and Time of Irradiation:	01-Aug-2014 22:37
Reference Dose Range kGy:	34.7 - 35.1
Calculated Minimum Dose kGy:	30.8
Calculated Maximum Dose kGy:	38.5

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM  
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069