

www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 08-Sep-2014

MY03S11233098-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

NITRITEX (M) SDN BHD
No.2, Jalan Jurunilai U1/20
Seksyen U1, Hicom Glenmarie
Industrial Park
40150 Shah Alam
Selangor

Order Information

Account Number:	101178								
Synergy Health Sales Part Reference:	1072642								
Customer Reference Number:	SNB14/0129								
Product Description:	BIOCLEAN ALPHA ST LTX (AL300)								
Quantity Received:	44								
Customer Minimum Specification kGy:	25.0								
Customer Maximum Specification kGy:	40.0								
Customer Unit Lot/Batch Number:	.								
Other Process Details:	DESCRIPTION: BIOCLEAN ALPHA ST LATEX (AL300) PACKING : 200 PAIRS/CTN								
	<table><tr><td>PRODUCT CODE</td><td>LOT NO</td><td>BATCH NO</td><td>QTY</td></tr><tr><td>AL300-70</td><td>14-1506</td><td>L31408024</td><td>44</td></tr></table>	PRODUCT CODE	LOT NO	BATCH NO	QTY	AL300-70	14-1506	L31408024	44
PRODUCT CODE	LOT NO	BATCH NO	QTY						
AL300-70	14-1506	L31408024	44						
	ACTUAL DOSE RECEIVED: MIN:34.3 kGy MAX:37.3 kGy								

Irradiation Data

Date and Time of Irradiation:	06-Sep-2014 11:57
-------------------------------	-------------------

Irradiation Release Authorised By Synergy Health plc

Registered Office: Suite 18.01 18th Floor, MWE Plaza, 8, Lebuh Farquhar, 10200, Penang, MALAYSIA

VAT Number:

Page 1 of 1