

http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 18-May-2015 UK33S11387543-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Order Information

Synergy Health Sales Part Reference:

Customer Reference Number:

Product Description:

Validation Reference:

Quantity Received:

Customer Minimum Specification kGy:

Customer Maximum Specification kGy:

Customer Unit Lot/Batch Number:

Other Process Details:

124123

1073638

0000007130

Latex Gloves Group 1 DV4172 25-45kGy

4172

60

25.0

45.0

0000007130

Irradiation No: S1138754311

Description: Bioclean Alpha St Latex (AL300)

Packing: 200 Pairs /ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

15-0460 L31502029 AL300-85 12

Description: Bioclean Advance St Latex (BASL)

Packing: 200 Pairs /ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

15-0705 L11503024 BASL6.5 48

Total: 60 Cartons

Irradiation Data

Date and Time of Irradiation: 17-May-2015 13:54



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 18-May-2015 UK33S11387543-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Reference Dose Range kGy:	33.9 - 34.7
Calculated Minimum Dose kGy:	30.1
Calculated Maximum Dose kGy:	38.1