

http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 22-Jun-2015 UK33S11405685-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Order Information

Account Number:	124123
Synergy Health Sales Part Reference:	1073638
Customer Reference Number:	0000007191
Product Description:	Latex Gloves Group 1 DV4172 25-45kGy
Validation Reference:	4172
Quantity Received:	57
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	0000007191
Other Process Details:	Irradiation No: S1140568511
	Description: Bioclean Alpha St Latex (AL300)
	Packing: 200 Pairs/ ctn
	Lot No Batch No Product Code Qty(ctns)
	/ Size
	15-0817 L31504006 AL300-75 48
	15-0818 L31504007 AL300-80 9

Irradiation Data

Date and Time of Irradiation:	22-Jun-2015 06:03
Reference Dose Range kGy:	32.8 - 33.2
Calculated Minimum Dose kGy:	29.1
Calculated Maximum Dose kGy:	36.4

Irradiation Release Authorised By Synergy Health plc