


<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 03-Aug-2014

UK33S11208554-5-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Niritex Ltd
Minton Enterprise Park
Oaks Drive
Newmarket
Suffolk
CB8 7YY

Order Information

Account Number:	124123								
Synergy Health Sales Part Reference:	1073641								
Customer Reference Number:	0000006538								
Product Description:	BIOCLEAN C&D GROUP 27 DV4166 25-38kGy								
Validation Reference:	4166								
Quantity Received:	34								
Customer Minimum Specification kGy:	25.0								
Customer Maximum Specification kGy:	38.0								
Customer Unit Lot/Batch Number:	0000006538								
Other Process Details:	Irradiation No: S1120855451 Description: Bioclean C Disp St. Sleevecover (S-BCSC) Packing: 180 Pieces/ ctn								
	<table><tr><th>Lot No</th><th>Batch No</th><th>Product Code / Size</th><th>Qty(ctns)</th></tr><tr><td>14-1068</td><td>-</td><td>S-BCSC</td><td>32</td></tr></table>	Lot No	Batch No	Product Code / Size	Qty(ctns)	14-1068	-	S-BCSC	32
Lot No	Batch No	Product Code / Size	Qty(ctns)						
14-1068	-	S-BCSC	32						
	Description: Bioclean D St. Lab Coat (S-BDLC) Packing: 30 Pieces/ ctn								
	<table><tr><th>Lot No</th><th>Batch No</th><th>Product Code / Size</th><th>Qty(ctns)</th></tr><tr><td>14-1039</td><td>-</td><td>S-BDLC-L</td><td>2</td></tr></table>	Lot No	Batch No	Product Code / Size	Qty(ctns)	14-1039	-	S-BDLC-L	2
Lot No	Batch No	Product Code / Size	Qty(ctns)						
14-1039	-	S-BDLC-L	2						
	Total: 34 Cartons								

Irradiation Data



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 03-Aug-2014

UK33S11208554-5-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd
Minton Enterprise Park
Oaks Drive
Newmarket
Suffolk
CB8 7YY

Date and Time of Irradiation:	03-Aug-2014 10:33
Reference Dose Range kGy:	32.3 - 32.6
Calculated Minimum Dose kGy:	29.6
Calculated Maximum Dose kGy:	34.8

Irradiation Release Authorised By Synergy Health plc

LM 05 Aug 14 Rev 2.

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069