

http://www.synergyhealthplc.com

## **Certificate of Irradiation**

Date Issued: 09-0ct-2014 UK33S11252136-4-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

## **Order Information**

**Account Number:** 124123 **Synergy Health Sales Part Reference:** 1073641 **Customer Reference Number:** 0000006699 **Product Description:** BIOCLEAN C&D GROUP 27 DV4166 25-38kGy Validation Reference: 4166 **Quantity Received:** 57 **Customer Minimum Specification kGy:** 25.0 **Customer Maximum Specification kGy:** 38.0 **Customer Unit Lot/Batch Number:** 000006699



http://www.synergyhealthplc.com

## **Certificate of Irradiation**

Date Issued: 09-0ct-2014 UK33S11252136-4-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Other Process Details:

Irradiation No: S1125213641

Description: B/Clean C St.Disp.Apron (S-BCDA-S)

Packing: 50 Pieces/ ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-1286 - S-BCDA-S 16

Description: Bioclean C Disp St Sleevecover

(S-BCSC)

Packing: 180 Pieces/ ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-1451 - S-BCSC 32

Description: Bioclean D St. Lab Coat (S-BDLC)

Packing: 30 Pieces/ ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-1557 - S-BDLC-L 2

Description: Bioclean D St. Lab Coat (S-BDLC)

Packing: 30 Pieces/ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-1556 - S-BDLC-M 3

Description: Bioclean D St. Lab Coat (S-BDLC)



http://www.synergyhealthplc.com

## **Certificate of Irradiation**

Date Issued: 09-0ct-2014 UK33S11252136-4-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Packing: 30 Pieces/ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-1558 - S-BDLC-XL 2

Description: Bioclean D St. Lab Coat (S-BDLC)

Packing: 30 Pieces/ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-1559 - S-BDLC-XXL 2

**Total: 57 Cartons** 

Irradiation Data	
Date and Time of Irradiation:	08-0ct-2014 13:31
Reference Dose Range kGy:	31.8 - 32.1
Calculated Minimum Dose kGy:	29.1
Calculated Maximum Dose kGy:	34.2