

http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 21-Apr-2015 UK33S11369317-8-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park **Oaks Drive** Newmarket Suffolk **CB8 7YY**

Order Information

Account	Numl	ber:
---------	------	------

Synergy Health Sales Part Reference:

Customer Reference Number:

Product Description:

Validation Reference:

Quantity Received:

Customer Minimum Specification kGy:

Customer Maximum Specification kGy:

Customer Unit Lot/Batch Number:

Other Process Details:

124123

1073641

0000007075

BIOCLEAN C&D GROUP 27 DV4166 25-38kGy

4166

148

25.0

38.0

0000007075

Irradiation No: S1136931781

Description: Bioclean C Disp St Sleevecover

(S-BCSC)

Packing: 180 Pieces/ctn

Lot No **Batch No** Product Code Qty(ctns)

/ Size

14-2361 S-BCSC 32 N/A

Description: Bioclean D St. Lab Coat (S-BDLC)

Packing: 30 Pieces/ctn

Lot No **Batch No** Product Code Qty(ctns) / Size 14-2136 N/A S-BDLC-L 8 108

15-0273 S-BDLC-L N/A

Total: 148 Cartons



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 21-Apr-2015 UK33S11369317-8-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Irradiation Data		
Date and Time of Irradiation:	21-Apr-2015 01:47	
Reference Dose Range kGy:	31.7 - 32.5	
Calculated Minimum Dose kGy:	29.0	
Calculated Maximum Dose kGy:	34.7	