



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 30-Jul-2015

UK33S11423995-8-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Niritex Ltd
 Minton Enterprise Park
 Oaks Drive
 Newmarket
 Suffolk
 CB8 7YY

Order Information

Account Number:	124123
Synergy Health Sales Part Reference:	1073641
Customer Reference Number:	0000007245
Product Description:	BIOCLEAN C&D GROUP 27 DV4166 25-38kGy
Validation Reference:	4166
Quantity Received:	117
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	38.0
Customer Unit Lot/Batch Number:	0000007245, 2pts



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 30-Jul-2015

UK33S11423995-8-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Niritex Ltd
 Minton Enterprise Park
 Oaks Drive
 Newmarket
 Suffolk
 CB8 7YY

Other Process Details:

Irradiation No: **S1142399581**
 Description: Bioclean C Disp St Sleevecover (S-BCSC)
 Packing: 180 Pairs /ctn

Lot No	Batch No	Product Code / Size	Qty(ctns)
15-1208	N/A	S-BCSC	32

Irradiation No: **S1142399581**
 Description: Bioclean D St Cvrll Collar (S-BDCCT)
 Packing: 20 Pairs /ctn

Lot No	Batch No	Product Code / Size	Qty(ctns)
15-0399	N/A	S-BDCCT-XL	1
15-0995	N/A	S-BDCCT-XL	17

Irradiation No: **S1142399581**
 Description: Bioclean D St Sleevecover (S-BDSC-L)
 Packing: 180 Pairs /ctn

Lot No	Batch No	Product Code / Size	Qty(ctns)
15-0929	N/A	S-BDSC-L	67

Total: 117 Cartons



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 30-Jul-2015

UK33S11423995-8-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

Niritex Ltd
 Minton Enterprise Park
 Oaks Drive
 Newmarket
 Suffolk
 CB8 7YY

Irradiation Data

Date and Time of Irradiation:	28-Jul-2015 21:19
Reference Dose Range kGy:	33.0 - 33.7
Calculated Minimum Dose kGy:	30.2
Calculated Maximum Dose kGy:	35.9

Irradiation Release Authorised By Synergy Health plc

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069