



www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 28-Jul-2015

MY01S11432303-1-1

This is to certify that Synergy Sterilisation Rawang (M) Sdn. Bhd. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

NITRITEX (M) SDN BHD
 No.2, Jalan Jurunilai U1/20
 Seksyen U1, Hicom Glenmarie
 Industrial Park
 40150 Shah Alam
 Selangor

Order Information

Account Number:	101178								
Synergy Health Sales Part Reference:	1097195								
Customer Reference Number:	SNB15/0056								
Product Description:	BIOCLEAN C DISP ST SLEEVECOVER (S-BCSC)								
Quantity Received:	48								
Customer Minimum Specification kGy:	25.0								
Customer Maximum Specification kGy:	38.0								
Other Process Details:	DESCRIPTION: BIOCLEAN C DISP ST SLEEVECOVER (S-BCSC) PACKING : 90 PAIRS/CTN								
	<table><tr><td>PRODUCT CODE</td><td>LOT NO</td><td>BATCH NO</td><td>QTY</td></tr><tr><td>S-BCSC</td><td>15-1708</td><td>-</td><td>48</td></tr></table>	PRODUCT CODE	LOT NO	BATCH NO	QTY	S-BCSC	15-1708	-	48
PRODUCT CODE	LOT NO	BATCH NO	QTY						
S-BCSC	15-1708	-	48						
	ACTUAL DOSE RECEIVED: MIN:29.1 kGy MAX:37.9 kGy								

Irradiation Data

Date and Time of Irradiation:	28-Jul-2015 10:28
-------------------------------	-------------------

Irradiation Release Authorised By Synergy Health plc