



www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 27-May-2014

MY03S11168806-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

NITRITEX (M) SDN BHD
 No.2, Jalan Jurunilai U1/20
 Seksyen U1, Hicom Glenmarie
 Industrial Park
 40150 Shah Alam
 Selangor

Order Information

Account Number:	101178										
Synergy Health Sales Part Reference:	1073188										
Customer Reference Number:	SNB14/0105										
Product Description:	B/CLEAN P/FLOW ST CR MARKER (S-BPFP)										
Quantity Received:	10										
Customer Minimum Specification kGy:	25.0										
Customer Maximum Specification kGy:	38.0										
Customer Unit Lot/Batch Number:	.										
Other Process Details:	DESCRIPTION: B/CLEAN P/FLOW ST CR MARKER (S-BPFP) PACKING:200 PIECES/CTN <table> <tr> <td>PRODUCT</td><td>LOT NO</td></tr> <tr> <td>CODE</td><td></td></tr> <tr> <td>S-BPFP</td><td>14-0909</td></tr> <tr> <td>QTY(CTN)</td><td></td></tr> <tr> <td>10</td><td></td></tr> </table> ACTUAL DOSE RECEIVED: MIN:30.6 kGy MAX:36.5 kGy	PRODUCT	LOT NO	CODE		S-BPFP	14-0909	QTY(CTN)		10	
PRODUCT	LOT NO										
CODE											
S-BPFP	14-0909										
QTY(CTN)											
10											

Irradiation Data

Date and Time of Irradiation:	27-May-2014 10:12
-------------------------------	-------------------

Registered Office: Suite 18.01 18th Floor, MWE Plaza, 8, Lebuhr Farquhar, 10200, Penang, MALAYSIA

VAT Number:



www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 27-May-2014

MY03S11168806-1-1

This is to certify that Synergy Sterilisation (M) Sdn. Bhd. has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products
EN ISO 9001:2008 Quality Management System
EN ISO 13485:2012 Quality System - Medical Devices

NITRITEK (M) SDN BHD
No.2, Jalan Jurunilai U1/20
Seksyen U1, Hicom Glenmarie
Industrial Park
40150 Shah Alam
Selangor

Irradiation Release Authorised By Synergy Health plc

Registered Office: Suite 18.01 18th Floor, MWE Plaza, 8, Lebuhr Farquhar, 10200, Penang, MALAYSIA

VAT Number:

Page 2 of 2