

http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 26-Sep-2014 UK33S11242642-7-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park **Oaks Drive** Newmarket Suffolk **CB8 7YY**

Order Information

Account Number: 124123 **Synergy Health Sales Part Reference:** 1076094 **Customer Reference Number:** 0000006665 **Product Description:**

B/CLEAN PERMAFLOW PENS VALIDATION

Quantity Received: Customer Minimum Specification kGy: 25.0 **Customer Maximum Specification kGy:** 45.0 Customer Unit Lot/Batch Number: 0000006665, VALIDATION



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 26-Sep-2014 UK33S11242642-7-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Other Process Details:

Irradiation No: S1124264271

Description: B/Clean P/flow ST CR Marker BK

(S-BPFP)

Packing: 200 Pieces/ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-1457 - S-BPFP-BK1 1

Description: B/Clean P/flow ST CR Marker BK

(S-BPFP)

Packing: 200 Pieces/ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-1454 - S-BPFP-RD1 1

Description: B/Clean P/flow ST CR Marker BK

(S-BPFP)

Packing: 200 Pieces/ctn

Lot No Batch No Product Code Qty(ctns)

/ Size

14-0909 - S-BPFP-BL1 2

Description: B/Clean P/flow ST CR Marker BK

(S-BPFP)

Packing: 200 Pieces/ctn

Lot No Batch No Product Code Qty(ctns)

/ Size



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 26-Sep-2014 UK33S11242642-7-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Date and Time of Irradiation:	25-Sep-2014 12:32
	Actual Min Recieved 29.6kGy Actual Max Recieved 36.2kGy
	Total: 6 Cartons
	14-0909 - S-BPFP-BK1 2

Irradiation Release Authorised By Synergy Health plc