

http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 08-Dec-2014 UK33S11288237-9-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Nitritex Ltd Minton Enterprise Park Oaks Drive Newmarket Suffolk CB8 7YY

Order Information

| | Total: 3 Cartons |
|--------------------------------------|---|
| | / Size 14-1999 - S-BPFP-BK1 3 |
| | Lot No Batch No Product Code Qty(ctns) |
| | Packing: 200 Pieces/ctn |
| | (S-BPFP) |
| Other Process Details: | Irradiation No: S1128823791 Description: B/Clean P/flow ST CR Marker BK |
| Customer Unit Lot/Batch Number: | 0000006806 |
| Customer Maximum Specification kGy: | 45.0 |
| Customer Minimum Specification kGy: | 25.0 |
| Quantity Received: | 3 |
| Validation Reference: | 4235 |
| Product Description: | B/CLEAN PERMAFLOW PENS DV4235 25-45kGy |
| Customer Reference Number: | 000006806 |
| Synergy Health Sales Part Reference: | 1076094 |
| Account Number: | 124123 |

Irradiation Release Authorised By Synergy Health plc

Date and Time of Irradiation:

Reference Dose Range kGy:

Calculated Minimum Dose kGy:

Calculated Maximum Dose kGy:

08-Dec-2014 01:21

32.5 - 32.5

28.5

36.9