



525 Locust Grove
Spartanburg, SC 29303

864-503-8333
1-800-289-5762

fax 864-503-8444
www.contecinc.com

CERTIFICATE OF CONFORMANCE AND STERILITY

Date: July 30, 2013

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

Product Code: MEQT0002
Lot Number: 297153-*-*1
Manufacturing Date: 07/19/2013
Run Number: 41167A
Irradiation Dose Specified: 25kGy to 50kGy
Irradiation Dose Delivered: 25.2kGy to 39.2kGy
Validated Sterile Product: Validation Sterile packets available upon request.
Sterility Assurance Level: 10^{-6} (SAL determined by test method
AAM1/ISO 11137)
Expiration Date: 07/2015

QC Supervisor
coc@contecinc.com

Certificate Of Processing

Prepared for CONTEC INC



Isomedix Services

Gamma Process Run ID 41167A

<u>Product Code</u>	<u>Product Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
33502-730	296142-*-*-1	24	CS
33502-730	296143-*-*-1	24	CS
B3340-4-IR	23946-S-CR2	41	CS
LWCS0027	294273-*-*-1	18	CS
LWCS0027	294274-*-*-1	18	CS
LWCS0027	294275-*-*-1	18	CS
LWPS0007	294742-*-*-1	20	CS
MEQT0002	297153-*-*-1	72	CS
MVK-MOP-IR	290950-*-*-1	18	CS
PN-99IR	295207-*-*-1	26	CS

Processing Run Start Date/Time:	29-Jul-2013 02:12:52 am	Approx. Downtime (hours):	0.00
Processing Run End Date/Time:	29-Jul-2013 12:41:47 pm		

Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	25.2
Maximum Specified Dose (kGy):	50.0	Maximum Delivered Dose (kGy):	39.2

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest



Reviewed and E-Signed By
Randy Miller (Production Planner)
Document Content Revision: 1

Signed On 7/29/2013 at 3:05 PM
UTC / GMT Offset (hh:mm): -4:00

Processing Location:
STERIS Isomedix Services
2072 Southport Road
Spartanburg, SC 29306
Phone: 864-582-3041
Fax: 864-582-4075

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2003/2012, and in alignment with the applicable standard, EN ANSI/AAMI/ISO 11137:2006 or EN ANSI/AAMI/ISO 11135:2007. For items processed with gamma irradiation, STERIS Isomedix certifies that these items received the indicated doses within the precision and accuracy of the dosimetry system used.