



525 Locust Grove  
Spartanburg, SC 29303

864-503-8333  
1-800-289-5762

fax 864-503-8444  
www.contecinc.com

**CERTIFICATE OF CONFORMANCE AND STERILITY**

Date: February 26, 2014

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

Product Code: MEQT0002  
Lot Number: 338000-\*-\*1  
Manufacturing Date: 02/20/2014  
Run Number: 42492A  
Irradiation Dose Specified: 25kGy to 50kGy  
Irradiation Dose Delivered: 26.8kGy to 41.6kGy  
Validated Sterile Product: Validation Sterile packets available upon request.  
Sterility Assurance Level:  $10^{-6}$  (SAL determined by test method  
AAM1/ISO 11137)  
Expiration Date: 02/2016

QC Supervisor  
coc@contecinc.com

# Certificate Of Processing

Prepared for CONTEC INC



Isomedix Services

Gamma Process Run ID 42492A

<u>Product Code</u>	<u>Product Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
21914-772	336429-**-1	48	CS
21914-772	336430-**-1	48	CS
21914-772	336431-**-1	48	CS
33502-730	336810-**-1	24	CS
ELMOP3IR	335616-**-1	48	CS
EL-MOP-IR	335587-**-1	18	CS
EZMP0150	331098-**-1	16	CS
LVK-MOP-IR	333132-**-1	18	CS
LVK-MOP-IR	335386-**-1	18	CS
LWCS0033	337168-**-1	36	CS
MEQT0002	338000-**-1	72	CS
MVK-MOP-IR	333147-**-1	18	CS
SB32IR	337209-**-1	32	CS

Processing Run Start Date/Time:	25-Feb-2014 01:51:01 pm	Approx. Downtime (hours):	0.00
Processing Run End Date/Time:	26-Feb-2014 04:39:00 am		

Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	26.8
Maximum Specified Dose (kGy):	50.0	Maximum Delivered Dose (kGy):	41.6

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

## Signature Manifest

Reviewed and E-Signed By

**Shawn Laurent (Production Supervisor)**

Document Content Revision: 1

Signed On 2/26/2014 at 4:41 AM

UTC / GMT Offset (hh:mm): -5:00

### Processing Location:

STERIS Isomedix Services  
2072 Southport Road  
Spartanburg, SC 29306  
Phone: 864-582-3041  
Fax: 864-582-4075

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2003/2012, and in alignment with the applicable standard, EN ANSI/AAMI/ISO 11137:2006 or EN ANSI/AAMI/ISO 11135:2007. For items processed with gamma irradiation, STERIS Isomedix certifies that these items received the indicated doses within the precision and accuracy of the dosimetry system used.