



525 Locust Grove
Spartanburg, SC 29303

864-503-8333
1-800-289-5762

fax 864-503-8444
www.contecinc.com

CERTIFICATE OF CONFORMANCE AND STERILITY

Date: December 01, 2014

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

Product Code: MEQT0002
Lot Number: 385986-*-*1
Manufacturing Date: 11/23/2014
Run Number: 44295A
Irradiation Dose Specified: 25kGy to 50kGy
Irradiation Dose Delivered: 25.9kGy to 42.3kGy
Validated Sterile Product: Validation Sterile packets available upon request.
Sterility Assurance Level: 10^{-6} (SAL determined by test method
AAM1/ISO 11137)
Expiration Date: 11/2016

QC Supervisor
coc@contecinc.com

Certificate Of Processing

Prepared for CONTEC INC



Gamma Process Run ID 44295A

<u>Product Code</u>	<u>Product Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
115-1724	382347-**-1	72	CS
33502-730	385310-**-1	24	CS
492226-944	384035-**-1	36	CS
LVK-MOP-IR	371617-**-1	18	CS
LWQT0031	380593-**-1	20	CS
MEQT0002	385986-**-1	72	CS
MVK-MOP-IR	382203-**-1	18	CS
NWPZ0001	130996-**-3	4	CS
NWPZ0001	132399-**-1	96	CS
TMVKMOPIR	383102-**-1	18	CS

Processing Run Start Date/Time:	28-Nov-2014 07:45:26 am	Approx. Downtime (hours):	3.40
Processing Run End Date/Time:	28-Nov-2014 09:44:01 pm		

Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	25.9
Maximum Specified Dose (kGy):	50.0	Maximum Delivered Dose (kGy):	42.3

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest

Reviewed and E-Signed By
Princess Shaw (QS/RC Technician)
Document Content Revision: 1

Signed On 12/1/2014 at 8:39 AM
UTC / GMT Offset (hh:mm): -5:00

Processing Location:
STERIS Isomedix Services
2072 Southport Road
Spartanburg, SC 29306
Phone: 864-582-3041
Fax: 864-582-4075

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2003/2012, and in alignment with the applicable standard, EN ANSI/AAMI/ISO 11137:2006 or EN ANSI/AAMI/ISO 11135:2007. For items processed with gamma irradiation, STERIS Isomedix certifies that these items received the indicated doses within the precision and accuracy of the dosimetry system used.