



525 Locust Grove
Spartanburg, SC 29303

864-503-8333
1-800-289-5762

fax 864-503-8444
www.contecinc.com

CERTIFICATE OF CONFORMANCE AND STERILITY

Date: July 21, 2015

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

Product Code: MEQT0002
Lot Number: 422689-*-*1
Manufacturing Date: 07/15/2015
Run Number: 45734A
Irradiation Dose Specified: 25.0kGy to 50.0kGy
Irradiation Dose Delivered: 27.5kGy to 45.6kGy
Validated Sterile Product: Validation Sterile packets available upon request.
Sterility Assurance Level: 10^{-6} (SAL determined by test method
AAMI/ISO 11137)
Expiration Date: 07/2017

QC Supervisor
coc@contecinc.com

Certificate Of Processing

Prepared for CONTEC INC



Isomedix Services

Gamma Process Run ID 45734A

<u>Product Code</u>	<u>Product Lot Number</u>	<u>Quantity</u>	<u>UOM</u>
21914-772	416333-**-1	48	CS
33502-730	425160-**-1	24	CS
492226-939	419387-**-1	36	CS
492226-944	418619-**-1	36	CS
B3340-4-IR	28665-S-CR2	87	CS
BLVKMOPIR	414853-**-1	18	CS
ELMOP3IR	422741-**-1	60	CS
EL-MOP-IR	422790-**-1	18	CS
LWPS0007	420186-**-1	20	CS
MEQT0002	420182-**-1	72	CS
MEQT0002	420183-**-1	72	CS
MEQT0002	422689-**-1	72	CS
MVK-MOP-IR	410529-**-1	18	CS
MVK-MOP-IR	410530-**-1	18	CS
NWPZ0001	136492-**-2	2	CS
NWPZ0001	136906-**-3	4	CS
NWPZ0001	136906-**-4	19	CS

Processing Run Start Date/Time: 19-Jul-2015 01:37:50 pm

Approx. Downtime (hours): 2.50

Processing Run End Date/Time: 20-Jul-2015 06:30:12 am

Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	27.5
Maximum Specified Dose (kGy):	50.0	Maximum Delivered Dose (kGy):	45.6

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

Signature Manifest

Reviewed and E-Signed By



Princess Shaw (QS/RC Technician)

Document Content Revision: 1

Signed On 7/20/2015 at 1:47 PM

UTC / GMT Offset (hh:mm): -4:00

Processing Location:

STERIS Isomedix Services
2072 Southport Road
Spartanburg, SC 29306
Phone: 864-582-3041
Fax: 864-582-4075

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2003/2012, and in alignment with the applicable standard, EN ANSI/AAMI/ISO 11137:2006 or EN ANSI/AAMI/ISO 11135:2007. For items processed with gamma irradiation, STERIS Isomedix certifies that these items received the indicated doses within the precision and accuracy of the dosimetry system used.