



525 Locust Grove  
Spartanburg, SC 29303

864-503-8333  
1-800-289-5762

fax 864-503-8444  
www.contecinc.com

## CERTIFICATE OF CONFORMANCE AND STERILITY

Date: August 11, 2015

This certification is provided as full assurance that the following product code and lot number was manufactured in accordance with prescribed procedures and specifications.

---

Product Code: MEQT0002  
Lot Number: 425248-\*-\*1  
Manufacturing Date: 08/03/2015  
Run Number: 45862A  
Irradiation Dose Specified: 25.0kGy to 50.0kGy  
Irradiation Dose Delivered: 31.2kGy to 45.8kGy  
Validated Sterile Product: Validation Sterile packets available upon request.  
Sterility Assurance Level:  $10^{-6}$  (SAL determined by test method  
AAMI/ISO 11137)  
Expiration Date: 08/2017

---

A handwritten signature in black ink, appearing to read "C. Duffee".

QC Supervisor  
coc@contecinc.com

# Certificate Of Processing

Prepared for CONTEC INC



Isomedix Services

Gamma Process Run ID 45862A

Product Code	Product Lot Number	Quantity	UOM
33502-730	429439-**-1	24	CS
33502-730	429444-**-1	24	CS
492238-960	417744-**-1	18	CS
EL-MOP-IR	429366-**-1	18	CS
LWCS0027	424782-**-1	18	CS
MEQT0002	425248-**-1	72	CS
MVK-MOP-IR	416447-**-1	18	CS
MVK-MOP-IR	418746-**-1	18	CS
MVK-MOP-IR	418747-**-1	18	CS
PN-99IR	422751-**-1	26	CS
PN-99IR	422752-**-1	26	CS
SB32IR	421716-**-1	48	CS

Processing Run Start Date/Time:	09-Aug-2015 12:11:35 am	Approx. Downtime (hours):	1.30
Processing Run End Date/Time:	09-Aug-2015 11:26:09 am		

Minimum Specified Dose (kGy):	25.0	Minimum Delivered Dose (kGy):	31.2
Maximum Specified Dose (kGy):	50.0	Maximum Delivered Dose (kGy):	45.8

Product meets Customer specifications; zero nonconformities occurred during this irradiation run.

## Signature Manifest

Reviewed and E-Signed By  
**Princess Shaw (QS/RC Technician)**  
Document Content Revision: 1

Signed On 8/10/2015 at 11:39 AM  
UTC / GMT Offset (hh:mm): -4:00

**Processing Location:**  
STERIS Isomedix Services  
2072 Southport Road  
Spartanburg, SC 29306  
Phone: 864-582-3041  
Fax: 864-582-4075

Operating facilities are in compliance with applicable state and federal regulations (FDA, NRC, EPA, and OSHA) and provide services under a quality system which meets the requirements of FDA QSR, EN/ISO 13485:2003/2012, and in alignment with the applicable standard, EN ANSI/AAMI/ISO 11137:2008 or EN ANSI/AAMI/ISO 11135:2007. For items processed with gamma irradiation, STERIS Isomedix certifies that these items received the indicated doses within the precision and accuracy of the dosimetry system used.